

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/085,072	03/01/2002	Bozidar Ferek-Petric	P-8158.02 DIV1	1422	
27581 MEDTRONIC,	7590 03/20/2007 INC.	· ·	EXAMINER		
710 MEDTRONIC PARK			OROPEZA, FRANCES P		
MINNEAPOLI	S, MN 55432-9924	•	. ART UNIT	PAPER NUMBER	
			3766	-	
			<u></u>		
			MAIL DATE	DELIVERY MODE	
•			03/20/2007	 PAPER 	

Please find below and/or attached an Office communication concerning this application or proceeding.

Advisory Action						
Before	the Filing	of an	Appeal	Brief		

Application No.	Applicant(s)		
10/085,072	FEREK-PETRIC, BOZIDAR		
Examiner	Art Unit		
Frances P. Oropeza	3766		

--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --THE REPLY FILED 20 February 2007 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE. 1. X The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods: a) \square The period for reply expires $\underline{3}$ months from the mailing date of the final rejection. b) The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection. Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f). Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). NOTICE OF APPEAL 2. The Notice of Appeal was filed on _____. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a). 3. X The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because (a) They raise new issues that would require further consideration and/or search (see NOTE below); (b) They raise the issue of new matter (see NOTE below): (c) They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or (d) They present additional claims without canceling a corresponding number of finally rejected claims. NOTE: See Continuation Sheet. (See 37 CFR 1.116 and 41.33(a)). 4. The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324). 5. Applicant's reply has overcome the following rejection(s): ___ 6. Newly proposed or amended claim(s) _____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s). 7. X For purposes of appeal, the proposed amendment(s): a) X will not be entered, or b) X will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended. The status of the claim(s) is (or will be) as follows: Claim(s) allowed: Claim(s) objected to: Claim(s) rejected: 35 and 37-39. Claim(s) withdrawn from consideration: _____ AFFIDAVIT OR OTHER EVIDENCE 8. The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e). 9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1). 10. The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached. REQUEST FOR RECONSIDERATION/OTHER 11. The request for reconsideration has been considered but does NOT place the application in condition for allowance because: 12. Note the attached Information Disclosure Statement(s). (PTO/SB/08) Paper No(s). 13. Other: ____. FRANCES P.ORDPEZA 3 Carl H. Loyso
CARLLAYNO

ACTING SPE, AU 3766

ART UNIT 3766

Continuation of 3. NOTE:

In the response filed 2/20/07, the Applicant added new claims 51-56 and amended the independent claim 35 to include the following limitations noted in quotations: sensing ventricular signals for "both ventricles of " a patient's heart, a means carried on the coronary sinus lead for "both" producing a signal... "and for delivering pacing pulses and sensing electrical cardiac activity", to provide an output "signal" in response... and a device responsive to the output "signal" for delivering.... These new claims and amendments introduce new issues and require a new search and further consideration.

